









Information for calibrating flow and flow rate measuring instruments

Please supply us with the following information so that we can calibrate your measuring instruments as smoothly and quickly as possible:

Instrument Type:	Serial No.:	
Sensor adaptation		
☐ Höntzsch standard		
☐ Special connection		
If the equipment to be calibrated is other technical accessories relevan	not from Höntzsch, please supply us with the operating instructions and any	у
Measured variables ISO		
☐ Flow velocity	☐ Flow rate	
☐ Mass flow	☐ Temperature	
Measured variables DAkkS		
☐ Flow velocity		
Application conditions		
☐ Gas Type:		
Pressure: Ten	perature: Measuring range:	
In health hazardous substances:		
☐ No ☐ Yes, Declaration	n of Decontamination necessary (Höntzsch ISO9001_F830-3)	
☐ ISO Calibration (6 values) DAkkS Calibration (6 values	;)
☐ Calibration "as found" before	adjustment or repair	
☐ Calibration "as left" after adju	stment or repair	
☐ Estimate before repair and ac	justment	
☐ Repair up to max	€ or% of replacement value	
Adjustment is carried out over the analog output	entire, instrument-specific measuring range, irrespective of the scaling of the	Э
6 calibration values spread	over the measuring range	
☐ equivalent to the scaling of th	analog output	
□ over the entire, instrument-sp	ecific measuring range, irrespective of the scaling of the analog output	
☐ individual values:		
Changing the parameter se	ttings	
☐ change analog output termin	al value to	
☐ change other parameters		